



Shri Mata Vaishno Devi University

Kakryal, Katra -182320, J&K

Department of Biotechnology

01991-285535, 285634, Ext: 2386, Fax: 01991-285692

Email: purchase.dobt@smvdu.ac.in/hod.biotech@smvdu.ac.in

Enquiry No. : SMVDU/DoBT/17/653-659

Date: 05.07.2017

“Enquiry Letter”

M/s _____

SUB: - Quotation for supply and Installation of Digital Magnetic Stirrer in the Department of Biotechnology of SMVDU

Dear Sir,

You are requested to submit your sealed quotation in favour of the undersigned for the following as per the terms and conditions given below and overleaf:

S.No.	Description	Qty Reqd.
1.	Instrument: Digital Magnetic Stirrer with hot plate Specifications: Annexure-I attached	2 Nos.

Terms & Conditions

1. Last date and time for receipt of quotation at Shri Mata Vaishno Devi University, Kakryal Katra - 182320 (J&K) is 15 Days from the date of Issue upto) 5:00 pm.(i.e. 20.07.2017)
2. Please submit your quotation in sealed cover superscribing the Quotation Enquiry number and due date failing which the quotation is liable to be rejected.
 - a. Price quoted for the material / equipment shall be final and valid for 60 days from the date of opening.
 - b. The material / equipment offered shall conform to relevant Indian / International standards.
3. Delivery of Material / equipment / stores mentioned in this quotation will be supplied by the supplier at SMVDU Kakryal, Katra-182320(J&K)
4. Rate offered must be for metric system of unit lengths or volume or weight. The items in the quotations. Wherever applicable, technical literature may please be enclosed alongwith quotation.
5. SMVDU reserves the right to increase the quantity mentioned in this enquiry and SMVDU reserves the right to split this quotation and place order on one or more suppliers/ quotationers and the right to reject partly or completely quotations without assigning any reasons thereof.
6. The quotation shall be complete in all respects and the firms shall give the make/Brand name offered.
7. The payment will be released after the receipt of all and complete material/equipment with complete accessories in good working condition.
8. Material/equipments shall be guaranteed from the minimum period of one year for satisfactory performance, workmanship and for the quality of material/equipment supplied by the firm from the date of delivery.
9. Quotation received after DUE DATE will not be considered. SMVDU shall not be responsible for delays in postal transit.
10. All taxes / duties should be clearly mentioned in the tender.
11. ELECTRICALS: All electrical equipments should be suitable for operation on 230V, 50 Hz or 440V, 50Hz, 3 phases as applicable and must be tropicalised for Indian conditions.
12. AFTER SALES SERVICE: The offer shall clearly state full details of the after sales service facility available for the equipment. The details shall include qualification of service personnel stationed at the service centre, response time in case emergency, availability of spares etc. The offer shall also give the amount and other terms & conditions for annual service contracts of the equipments after the expiry of the guarantee / warrantee period.

13. **PACKING:**
 - a. The equipment shall be packed suitable for dispatch direct to SMVDU at supplier expenses and supplier shall be responsible for any damage during transit.
14. **JURISDICTIONS FOR DISPUTES:** In all disputes the decision of the Registrar , SMVDU, shall be final , conclusive and binding on the supplier / . All disputes shall be subject to the jurisdiction of the courts in the District Reasi.
15. **DEVIATIONS:** Deviations if any from the specifications given which provide for improvement in the functioning of the equipments will be accepted. Such deviations and their advantages shall be clearly brought out in the tender

Head of Department of Biotechnology
Head

Department of Biotechnology
SMVD University, Katra

ANNEXURE - I

Technical Specifications for Digital Magnetic Stirrer

- Number of stirring position one.
- Maximum stirring quantity (H_2O) should have 20 l.
- Speed range should have 50 - 1500 rpm.
- Should have temperature range: RT - 300 °C
- Should have integrated temperature control.
- Includes PT 1000 temperature sensor (PT 1000.60).
- Should display exact temperature and speed setting via digital display, even when switched off.
- Motor rating output should be 9 W.
- Heat output should be 600 W.
- Should have setting accuracy of speed 10 RPM or better
- Set safety temperature limit displayed digitally.
- Should have hot top indicator, hot surface warning to prevent burns.
- Should have adjustable safety circuit of heating plate temperature (50 - 360 °C)
- Should have high level of safety to improved heat control technology
- Should have enclosed assembly (IP 42) guarantees long service life.
- Should have aluminum heating plate of 135mm for optimum heat transfer
- Should have protection cover H 100.
- The system should have European CE certified.